<picture>
Spectrometers: Bruker IFS 66v , Bruker IFS 66v/S

**Parameters**

Range (cm$^{-1}$) : 10 - 70

D* (cm$^{-1}$/W/Hz$^{-1/2}$) 2x10$^{8}$

Time constant (s) : 10$^{-2}$

Operating temperature (K) RT

Suggested mirror velocity (HeNe modulation freq, kHz) : 3 - 5

<spectra>
<link to>How to change a detector?